

UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Addease COMMISSIONER FOR PATENTS PO Box 1430 Alexandria, Virginia 22313-1450 www.wopto.gov

| APPLICATION NO. | FILING DATE | FIRST NAMED INVENTOR | ATTORNEY DOCKET NO. | CONFIRMATION NO. |
|---|-------------|----------------------|---------------------|------------------|
| 10/553,336 | 11/27/2006 | Kenji Nagakawa | 10921.363USWO | 1215 |
| 52835 7550 90690/2009 HAMRE, SCHUMANN, MUELLER & LARSON, P.C. P.O. BOX 2902 MINNEAPOLIS, MN 55402-0902 | | | EXAMINER | |
| | | | SAKELARIS, SALLY A | |
| | | | ART UNIT | PAPER NUMBER |
| | | | 1797 | |
| | | | | |
| | | | MAIL DATE | DELIVERY MODE |
| | | | 06/09/2009 | PAPER |

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Application No. Applicant(s) 10/553,336 NAGAKAWA ET AL. Office Action Summary Examiner Art Unit Sally A. Sakelaris 1797 -- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --Period for Reply A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 1 MONTH(S) OR THIRTY (30) DAYS. WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION. Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication. If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication - Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b). Status 1) Responsive to communication(s) filed on 11 December 2008. 2a) This action is FINAL. 2b) This action is non-final. 3) Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under Ex parte Quayle, 1935 C.D. 11, 453 O.G. 213. Disposition of Claims 4) Claim(s) 1-24 is/are pending in the application. 4a) Of the above claim(s) is/are withdrawn from consideration. 5) Claim(s) _____ is/are allowed. 6) Claim(s) _____ is/are rejected 7) Claim(s) is/are objected to. 8) Claim(s) 1-24 are subject to restriction and/or election requirement. Application Papers 9) The specification is objected to by the Examiner. 10) The drawing(s) filed on is/are; a) accepted or b) objected to by the Examiner. Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a). Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d). 11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152. Priority under 35 U.S.C. § 119 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some * c) None of: Certified copies of the priority documents have been received. 2. Certified copies of the priority documents have been received in Application No. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)). * See the attached detailed Office action for a list of the certified copies not received. Attachment(s)

1) Notice of References Cited (PTO-892)

Notice of Draftsperson's Patent Drawing Review (PTO-948)

Information Disclosure Statement(s) (FTO/S5/08)
 Paper No(s)/Mail Date _______.

Interview Summary (PTO-413)
 Paper No(s)/Mail Date.

6) Other:

5 Notice of Informal Patent Application

Page 2

Application/Control Number: 10/553,336

Art Unit: 1797

Election/Restrictions

Restriction is required under 35 U.S.C. 121 and 372.

This application contains the following inventions or groups of inventions which are not so linked as to form a single general inventive concept under PCT Rule 13.1.

In accordance with 37 CFR 1.499, applicant is required, in reply to this action, to elect a single invention to which the claims must be restricted.

Group I, claim(s) 1-10 and 16-24 are drawn to an analyzing tool classified in for example, Class 422 subclass 55.

Group II, claim(s) 11-15, drawn to a method of making an analyzing tool as classified in for example, Class 436 subclass 164.

1. The inventions listed as Groups I-II lack unity of invention because even though the inventions of these groups require the technical feature of having a reaction space in which a particular component of a sample and a reagent react with each other; and a reagent portion which is arranged in the reaction space and which dissolves when the sample is supplied to the reaction space; wherein the reagent portion includes a first part and a second part facing each other and provided on a defining surface defining the reaction space, this technical feature is not a special technical feature as it does not make a contribution over the prior art in view of Watanabe et al. US Patent 2002/0134676 A1.

Watanabe et al. teach a highly responsive biosensor comprising a cholesterol-oxidizing enzyme and cholesterol esterase. The biosensor comprises: an electrically insulating base plate; an electrode system including a working electrode and a counter electrode formed on the base plate; a cover member joined with the base plate to define a sample supply pathway through which a sample solution is introduced from a sample supply port to the electrode system, the sample supply pathway being formed between the cover member and the base plate; and a reagent layer formed in the sample supply pathway, wherein the reagent layer contains a cholesterol-oxidizing enzyme, cholesterol esterase, an electron mediator, and a buffer having a buffering capacity in an acidic pH range (Abstract and Fig. 1).

The shared technical feature of Groups I-II above is therefore not a special technical feature as it does not make a contribution over the prior art.

The inventions are distinct, each from the other because of the following reasons:

Art Unit: 1797

2. Inventions I and II Inventions are related as process of making and product made. The inventions are distinct if either or both of the following can be shown: (1) that the process as claimed can be used to make another and materially different product or (2) that the product as claimed can be made by another and materially different process (MPEP § 806.05(f)). In the instant case the product as made could be produced by using the method disclosed in Watanabe of producing their biosensor that does not require the step of mounting the space to be performed before the intermediate product forming step.

Applicant is advised that the reply to this requirement to be complete must include (i) an
election of a species or invention to be examined even though the requirement may be traversed
(37 CFR 1.143) and (ii) identification of the claims encompassing the elected invention.

The election of an invention or species may be made with or without traverse. To preserve a right to petition, the election must be made with traverse. If the reply does not distinctly and specifically point out supposed errors in the restriction requirement, the election shall be treated as an election without traverse.

- 4. Applicant is reminded that upon the cancellation of claims to a non-elected invention, the inventorship must be amended in compliance with 37 CFR 1.48(b) if one or more of the currently named inventors is no longer an inventor of at least one claim remaining in the application. Any amendment of inventorship must be accompanied by a request under 37 CFR 1.48(b) and by the fee required under 37 CFR 1.17(i).
- 5. The examiner has required restriction between product and process claims. Where applicant elects claims directed to the product, and the product claims are subsequently found allowable, withdrawn process claims that depend from or otherwise require all the limitations of

Application/Control Number: 10/553,336

Art Unit: 1797

the allowable product claim will be considered for rejoinder. <u>All</u> claims directed to a nonelected process invention must require all the limitations of an allowable product claim for that process invention to be rejoined.

In the event of rejoinder, the requirement for restriction between the product claims and the rejoined process claims will be withdrawn, and the rejoined process claims will be fully examined for patentability in accordance with 37 CFR 1.104. Thus, to be allowable, the rejoined claims must meet all criteria for patentability including the requirements of 35 U.S.C. 101, 102, 103 and 112. Until all claims to the elected product are found allowable, an otherwise proper restriction requirement between product claims and process claims may be maintained. Withdrawn process claims that are not commensurate in scope with an allowable product claim will not be rejoined. See MPEP § 821.04(b). Additionally, in order to retain the right to rejoinder in accordance with the above policy, applicant is advised that the process claims should be amended during prosecution to require the limitations of the product claims. Failure to do so may result in a loss of the right to rejoinder. Further, note that the prohibition against double patenting rejections of 35 U.S.C. 121 does not apply where the restriction requirement is withdrawn by the examiner before the patent issues. See MPEP § 804.01.

 A telephone call was made to Doug Mueller on 5/27/2009 to request an oral election to the above restriction requirement, but did not result in an election being made.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Sally A. Sakelaris whose telephone number is 5712726297. The examiner can normally be reached on Monday-Friday 8-5:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Jill Warden can be reached on 5712721267. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Application/Control Number: 10/553,336 Page 5

Art Unit: 1797

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Sally Sakelaris /Jill Warden/

Supervisory Patent Examiner, Art Unit 1797